

PRODUCT RECALL

Date:	8/15/2024
Manufacturer:	Alcon
Article Numbers/ Product:	<ul style="list-style-type: none">• DAILIES TOTAL1® for Astigmatism• DAILIES TOTAL1® Multifocal• TOTAL30® Sphere contact lenses
Dates Sold:	6/5/2024 through 8/15/2024



Post Until: 2/15/2025

ECP Communication Package

Alcon Limited and Specific Voluntary Recall



Alcon Contact Lens Voluntary Recall

Dear Valued Customer,

Alcon has initiated a Voluntary Medical Device Recall for specific lots of DAILIES TOTAL1[®] for Astigmatism, DAILIES TOTAL1[®] Multifocal, and TOTAL30[®] Sphere soft contact lenses. Alcon identified an isolated quality issue with a material supplied by a third-party vendor that was used in the production of these specific contact lens lots at a single manufacturing site. As a result, the identified contact lens lots may not meet Alcon's quality and/or performance standards for the entirety of their labeled shelf life.

The contact lenses from the identified lots are not expected to cause an increased risk to the wearer. As with any contact lens, there is a remote likelihood that wearing a lens from the identified lots may lead to temporary issues such as irritation, redness, or blurred vision. These symptoms typically resolve after lens removal and/or replacement (as directed in the patient leaflet¹).

According to our records, affected contact lens lots may have been shipped to you. For your convenience, you can confirm if a lot of contact lenses is within the scope of the medical device recall at www.myalcon.com/search-my-lot/ or via the QR code below.

If your contact lenses are included in this medical device recall, please discontinue use and contact the dedicated support team for this voluntary recall at 1-888-266-7974 to assist you in arranging replacement contact lenses.

Thank you.



Scan to verify if your contact lenses
are included in the Voluntary
Medical Device Recall

1 – If you do not have a patient leaflet available, please visit ifu.alcon.com. Select “Contact Lenses” and search for your lens brand.